

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 013**

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024** to **01-10-2027**

Replaces annex dated: **20-07-2023**

Location(s) where activities are performed under accreditation

Head Office

Brainport Industries Campus (BIC) 1
 5657 BX
 Eindhoven
 The Netherlands

Location	Abbreviation/ location code
Brainport Industries Campus (BIC) 1 5657 BX Eindhoven The Netherlands	BIC
On-site at the customer	OS

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 0 0	DC/LF Quantities				
LF 1 0	DC Voltage			measuring and generating	BIC
	10 V		$5 \cdot 10^{-7} \cdot U$	zenerreferences	
	1 V and 1.018 V		$2.3 \cdot 10^{-6} \cdot U$	zenerreferences	
	0 µV to 10 µV		0.3 µV		
	10 µV to 200 mV		$3 \cdot 10^{-6} \cdot U + 0.2 \mu V$		
	200 mV to 1 V		$3 \cdot 10^{-6} \cdot U$		

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, U , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 013**

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024 to 01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
	1 V to 2 V		$2 \cdot 10^{-6} \cdot U$		
	2 V to 10 V		$1 \cdot 10^{-6} \cdot U$		
	10 V to 1000 V		$2 \cdot 10^{-6} \cdot U$		
	1000 V to 1100 V		$4 \cdot 10^{-6} \cdot U$		
LF 2 0	DC Current			measuring and generating	BIC
	1 μ A to 10 μ A		$3 \cdot 10^{-5} \cdot I$		
	10 μ A to 20 A		$2 \cdot 10^{-5} \cdot I$		
LF 3 0	AC Voltage			measuring and generating	BIC
	100 mV to 220 mV	10 Hz to 20 Hz	$3 \cdot 10^{-4} \cdot U$		
		20 Hz to 40 Hz	$5 \cdot 10^{-5} \cdot U$		
		40 Hz to 20 kHz	$5 \cdot 10^{-5} \cdot U$		
		20 kHz to 50 kHz	$5 \cdot 10^{-5} \cdot U$		
		50 kHz to 100 kHz	$6 \cdot 10^{-5} \cdot U$		
		100 kHz to 200 kHz	$2 \cdot 10^{-4} \cdot U$		
		200 kHz to 500 kHz	$4 \cdot 10^{-4} \cdot U$		
		500 kHz to 1 MHz	$7 \cdot 10^{-4} \cdot U$		
	220 mV to 2.2 V	10 Hz to 20 Hz	$5 \cdot 10^{-5} \cdot U$		
		20 Hz to 40 Hz	$5 \cdot 10^{-5} \cdot U$		
		40 Hz to 20 kHz	$5 \cdot 10^{-5} \cdot U$		
		20 kHz to 50 kHz	$4 \cdot 10^{-5} \cdot U$		
		50 kHz to 100 kHz	$5 \cdot 10^{-5} \cdot U$		
		100 kHz to 200 kHz	$2 \cdot 10^{-4} \cdot U$		
		200 kHz to 500 kHz	$4 \cdot 10^{-4} \cdot U$		
		500 kHz to 1 MHz	$7 \cdot 10^{-4} \cdot U$		
	2.2 V to 22 V	10 Hz to 20 Hz	$5 \cdot 10^{-5} \cdot U$		
		20 Hz to 40 Hz	$4 \cdot 10^{-5} \cdot U$		
		40 Hz to 20 kHz	$4 \cdot 10^{-5} \cdot U$		

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 013**

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024 to 01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
		20 kHz to 50 kHz	$4 \cdot 10^{-5} \cdot U$		
		50 kHz to 100 kHz	$5 \cdot 10^{-5} \cdot U$		
	2.2 V to 22 V	100 kHz to 200 kHz	$2 \cdot 10^{-4} \cdot U$		
		200 kHz to 500 kHz	$4 \cdot 10^{-4} \cdot U$		
		500 kHz to 1 MHz	$8 \cdot 10^{-4} \cdot U$		
	22 V to 220 V	10 Hz to 20 Hz	$6 \cdot 10^{-5} \cdot U$		
		20 Hz to 40 Hz	$5 \cdot 10^{-5} \cdot U$		
		40 Hz to 20 kHz	$5 \cdot 10^{-5} \cdot U$		
		20 kHz to 50 kHz	$5 \cdot 10^{-5} \cdot U$		
		50 kHz to 100 kHz	$2 \cdot 10^{-4} \cdot U$		
	220 V to 1000 V	10 Hz to 20 Hz	$5 \cdot 10^{-5} \cdot U$		
		20 Hz to 40 Hz	$5 \cdot 10^{-5} \cdot U$		
		40 Hz to 20 kHz	$5 \cdot 10^{-5} \cdot U$		
		20 kHz to 50 kHz	$7 \cdot 10^{-5} \cdot U$		
		50 kHz to 100 kHz	$4 \cdot 10^{-4} \cdot U$		
LF 3 3	Pulse Amplitude				BIC
	1 mV to 25 mV	10 Hz to 10 kHz	$5 \cdot 10^{-3} \cdot U$	measuring	
	25 mV to 110 mV	10 Hz to 10 kHz	$2.6 \cdot 10^{-4} \cdot U$		
	110 mV to 2.2 V	10 Hz to 10 kHz	$2.6 \cdot 10^{-4} \cdot U$		
	2.2 V to 11 V	10 Hz to 10 kHz	$2.6 \cdot 10^{-4} \cdot U$		
	11 V to 130 V	10 Hz to 10 kHz	$2.6 \cdot 10^{-4} \cdot U$		
	6 mV to 25 mV	10 Hz to 10 kHz	$1 \cdot 10^{-2} \cdot U$	generating	
	25 mV to 110 mV	10 Hz to 10 kHz	$5 \cdot 10^{-3} \cdot U$		
	110 mV to 2.2 V	10 Hz to 10 kHz	$5 \cdot 10^{-3} \cdot U$		
	2.2 V to 11 V	10 Hz to 10 kHz	$5 \cdot 10^{-3} \cdot U$		
	11 V to 130 V	10 Hz to 10 kHz	$5 \cdot 10^{-3} \cdot U$		

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 013**

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024 to 01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 3 4	AC/DC Transfer			measuring and generating	BIC
	0.5 V to 50 V	40 Hz to 1 kHz	$5.6 \cdot 10^{-5} \cdot U$		
		1 kHz to 20 kHz	$5.1 \cdot 10^{-5} \cdot U$		
		20 kHz to 100 kHz	$6.1 \cdot 10^{-5} \cdot U$		
		100 kHz to 500 kHz	$2.5 \cdot 10^{-4} \cdot U$		
	0.5 V to 10 V	500 kHz to 1 MHz	$4.2 \cdot 10^{-4} \cdot U$		
	50 V to 100 V	40 Hz to 1 kHz	$4.5 \cdot 10^{-5} \cdot U$		
		1 kHz to 20 kHz	$4.5 \cdot 10^{-5} \cdot U$		
		20 kHz to 50 kHz	$5.1 \cdot 10^{-5} \cdot U$		
		50 kHz to 100 kHz	$5.6 \cdot 10^{-5} \cdot U$		
	100 V to 500 V	40 Hz to 1 kHz	$6.2 \cdot 10^{-5} \cdot U$		
		1 kHz to 20 kHz	$5.8 \cdot 10^{-5} \cdot U$		
		20 kHz to 50 kHz	$9.2 \cdot 10^{-5} \cdot U$		
		50 kHz to 100 kHz	$2.4 \cdot 10^{-4} \cdot U$		
	500 V to 1000 V	40 Hz to 20 kHz	$6.4 \cdot 10^{-5} \cdot U$		
		20 kHz to 50 kHz	$9.6 \cdot 10^{-5} \cdot U$		
		50 kHz to 100 kHz	$2.4 \cdot 10^{-4} \cdot U$		
LF 4 0	AC Current			measuring and generating	BIC
	100 μ A to 1 mA	10 Hz to 1 kHz	$3.2 \cdot 10^{-4} \cdot I$		
		1 kHz to 5 kHz	$2.6 \cdot 10^{-4} \cdot I$		
		5 kHz to 10 kHz	$6.5 \cdot 10^{-4} \cdot I$		
		10 kHz to 20 kHz	$1.2 \cdot 10^{-3} \cdot I$	measuring only	
	1 mA to 10 mA	10 Hz to 1 kHz	$2.3 \cdot 10^{-4} \cdot I$		
		1 kHz to 5 kHz	$1.7 \cdot 10^{-4} \cdot I$		
		5 kHz to 10 kHz	$4.3 \cdot 10^{-4} \cdot I$		
		10 kHz to 20 kHz	$6.7 \cdot 10^{-4} \cdot I$	measuring only	

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 013**

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024 to 01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
	10 mA to 1 A	10 Hz to 1 kHz	$2.4 \cdot 10^{-4} \cdot I$		
		1 kHz to 5 kHz	$2.1 \cdot 10^{-4} \cdot I$		
		5 kHz to 10 kHz	$4.9 \cdot 10^{-4} \cdot I$		
		10 kHz to 20 kHz	$8.2 \cdot 10^{-4} \cdot I$	measuring only	
	1 A to 5 A	10 Hz to 1 kHz	$2.4 \cdot 10^{-4} \cdot I$		
		1 kHz to 5 kHz	$2.8 \cdot 10^{-4} \cdot I$		
		5 kHz to 10 kHz	$7.4 \cdot 10^{-4} \cdot I$		
		10 kHz to 20 kHz	$1.4 \cdot 10^{-3} \cdot I$	measuring only	
	5 A to 20 A	10 Hz to 1 kHz	$3.3 \cdot 10^{-4} \cdot I$		
		1 kHz to 5 kHz	$3.8 \cdot 10^{-4} \cdot I$		
		5 kHz to 10 kHz	$7.8 \cdot 10^{-4} \cdot I$		
		10 kHz to 20 kHz	$1.4 \cdot 10^{-3} \cdot I$	measuring only	
LF 6 2	DC Resistance			measuring and generating	BIC
	1 mΩ		$3 \cdot 10^{-5} \cdot R$		
	10 mΩ		$2 \cdot 10^{-5} \cdot R$		
	100 mΩ		$1 \cdot 10^{-5} \cdot R$		
	1 Ω		$3 \cdot 10^{-6} \cdot R$		
	10 Ω		$3 \cdot 10^{-6} \cdot R$		
	100 Ω		$3 \cdot 10^{-6} \cdot R$		
	1 kΩ		$3 \cdot 10^{-6} \cdot R$		
	10 kΩ		$2 \cdot 10^{-6} \cdot R$		
	100 kΩ		$3 \cdot 10^{-6} \cdot R$		
	1 MΩ		$3 \cdot 10^{-6} \cdot R$		
	10 MΩ		$5 \cdot 10^{-6} \cdot R$		
	100 MΩ		$2 \cdot 10^{-5} \cdot R$		
	1 GΩ		$6 \cdot 10^{-4} \cdot R$		
	1 mΩ to 10 mΩ		$9 \cdot 10^{-5} \cdot R$		
	10 mΩ to 100 mΩ		$3 \cdot 10^{-5} \cdot R$		

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 013**

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024 to 01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
	100 mΩ to 1 Ω		$2 \cdot 10^{-5} \cdot R$		
	1 Ω to 10 MΩ		$5 \cdot 10^{-6} \cdot R$		
	10 MΩ to 100 MΩ		$2 \cdot 10^{-5} \cdot R$		
	100 MΩ to 1 GΩ		$6 \cdot 10^{-4} \cdot R$		
LF 6 5	LF Capacitance			measuring and generating	BIC
	10 pF	1 kHz to 10 kHz	$2 \cdot 10^{-5} \cdot C$		
	100 pF	1 kHz to 10 kHz	$2 \cdot 10^{-5} \cdot C$		
	1000 pF	1 kHz to 10 kHz	$2 \cdot 10^{-5} \cdot C$		
	1 pF to 50 pF	1 kHz to 10 kHz	$2 \cdot 10^{-4} \cdot C$		
	50 pF to 100 pF	1 kHz to 10 kHz	$4 \cdot 10^{-5} \cdot C$		
	100 pF to 1 nF	1 kHz to 5 kHz	$2 \cdot 10^{-5} \cdot C$		
	1 nF to 1 μF	1 kHz	$2 \cdot 10^{-4} \cdot C$		
	1 μF to 10 μF	100 Hz	$3 \cdot 10^{-4} \cdot C$		
	200 μF to 500 μF	DCV	$1.2 \cdot 10^{-3} \cdot C$		
	500 μF to 110 mF	DCV	$1 \cdot 10^{-3} \cdot C$		
RF 0 0	High Frequency Quantities				
RF 1 0	CW Flatness				BIC
	5 mV _{pp} to 200 mV _{pp}	50 kHz to 1100 MHz	$3.5 \cdot 10^{-2}$ related to 50 kHz/50 Ω	measuring	
	200 mV _{pp} to 6 V _{pp}	50 kHz to 1100 MHz	$3.5 \cdot 10^{-2}$ related to 50 kHz/50 Ω	measuring	
	5 mV _{pp} to 20 mV _{pp}	50 kHz to 1100 MHz	$11 \cdot 10^{-2}$ related to 50 kHz/50 Ω	generating VSWR scope ≤ 1.3	
	20 mV _{pp} to 6 V _{pp}	50 kHz to 1100 MHz	$10 \cdot 10^{-2}$ related to 50 kHz/50 Ω	generating VSWR scope ≤ 1.3	

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024** to **01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
TF 0 0	Time and frequency				
TF 2 1	Frequency			measuring and generating	BIC
	10 MHz		$6 \cdot 10^{-11} \cdot f$		
	10 mHz to 1 MHz		$1 \cdot 10^{-10} \cdot f + T_e$	1)	
	1 MHz to 300 MHz		$1 \cdot 10^{-10} \cdot f$		
	300 MHz to 1.1 GHz		$6 \cdot 10^{-9} \cdot f$	2)	
TF 2 2	Time Interval				BIC
	1 μ s to 10 s		$5 \cdot 10^{-10} \cdot t + T.E.$	3) measuring only	
	10 s to 10 ⁵ s		$5 \cdot 10^{-10} \cdot t + 10$ ns	measuring only	
TF 2 3	Phase Angle				BIC
	0 ° to 180 °	10 Hz to 50 Hz	0.05 °	at equal input voltages 100 mV < U _i < 300 V generate up to 120 V	
		50 Hz to 1 kHz	0.08 °		
		1 kHz to 5 kHz	0.18 °		
		5 kHz to 10 kHz	0.35 °		
		10 kHz to 30 kHz	0.75 °		
		50 Hz	0.10 °	unequal input voltages 100 mV < U _i < 300 V ratio 1:100	
		50 Hz to 1 kHz	0.25 °		
		1 kHz to 5 kHz	0.40 °		
		5 kHz to 10 kHz	1.0 °		
		10 kHz to 30 kHz	1.8 °		
TF 2 4	Rise time				BIC
	70 ps to 1000 ps	pulse repeat ≤1 MHz	20 ps	180 mV _{pp} to 300 mV _{pp} in 50 Ω measuring only	

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024** to **01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ²	Remarks	Location
PV 0 0	Pressure				
PV 1 1	Absolute pressure	10 kPa to 60 kPa	12 Pa	nitrogen, measuring and generating	BIC
		60 kPa to 700 kPa	$3.1 \cdot 10^{-5} \cdot p + 10 \text{ Pa}$		
		700 kPa to 10 MPa	$4.2 \cdot 10^{-5} \cdot p + 95 \text{ Pa}$		
PV 1 2	Gauge pressure	0 kPa to 15 kPa	$2.5 \cdot 10^{-5} \cdot p_e + 0.7 \text{ Pa}$	nitrogen, measuring and generating	BIC
		15 kPa to 700 kPa	$3.5 \cdot 10^{-5} \cdot p_e + 10 \text{ Pa}$		
		700 kPa to 10 MPa	$4.2 \cdot 10^{-5} \cdot p_e + 95 \text{ Pa}$		
PV 3 1	Negative Gauge pressure	-90 kPa to -15 kPa	$3.5 \cdot 10^{-5} \cdot p_e + 6.1 \text{ Pa}$	nitrogen, measuring and generating	BIC
		-15 kPa to 0 kPa	$3.5 \cdot 10^{-5} \cdot p_e + 0.7 \text{ Pa}$		
TQ 0 0	Torque	0.45 to 5.6 Nm	$2 \cdot 10^{-2} \cdot M + 0.069 \text{ Nm}$	Setting Torque Tools (wrenches and drivers)	BIC
		5.6 to 41 Nm	$1.5 \cdot 10^{-2} \cdot M + 0.53 \text{ Nm}$		
		41 to 113 Nm	$7 \cdot 10^{-3} \cdot M + 0.86 \text{ Nm}$		
		113 to 339 Nm	$6 \cdot 10^{-3} \cdot M + 1.1 \text{ Nm}$		

² Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, U , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024 to 01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ²	Remarks	Location
TE 0 0	Temperature, Humidity, Thermophysical properties				
TE 1 0	Resistance thermometers	5 °C to 15 °C	0.11 °C	measurements in climate chamber	BIC
		15 °C to 24 °C	0.045 °C	measurements in climate chamber	BIC
		24 °C to 65 °C	0.045 °C to 0.17 °C	measurements in climate chamber	BIC
		-196 °C	0.014 °C	Liquid Nitrogen	BIC
		0.01 °C	0.0059 °C	triple point of water	BIC
		29.7646 °C	0.0068 °C	fixed point gallium	BIC
		419.527 °C	0.010 °C	fixed point zinc	BIC
		660.323 °C	0.014 °C	fixed point aluminium	BIC
		-95 °C to -80 °C	0.025 °C		BIC
		-80 °C to 248 °C	0.014 °C		BIC
		248 °C to 500 °C	0.021 °C		BIC
		500 °C to 660 °C	0.053 °C		BIC
TE 1 3	Thermistors	0 °C to 100 °C	0.014 °C		BIC
TE 3 0	Thermocouples	0 °C to 26 °C	0.022 °C	thermocouple Type-E	BIC
		-30 °C to 200 °C	0.16 °C		OS
		-95 °C to 660 °C	0.10 °C		BIC
		660 °C to 1000 °C	0.80 °C		BIC
		35 °C to 500 °C	0.5 °C to 1.8 °C	surface thermometers and surface calibrators	BIC

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024 to 01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ²	Remarks	Location
TE 4 0	Self-indicating thermometers				
TE 4 1	Temperature sensors with display unit (digital system thermometer or dataloggers)	5 °C to 15 °C	0.11 °C	measurements in climate chamber	BIC
		15 °C to 24 °C	0.045 °C	measurements in climate chamber	BIC
		24 °C to 65 °C	0.045 °C to 0.17 °C	measurements in climate chamber	BIC
		0.01 °C	0.0059 °C	triple point of water	BIC
		29.7646 °C	0.0068 °C	fixed point gallium	BIC
		419.527 °C	0.010 °C	fixed point zinc	BIC
		660.323 °C	0.014 °C	fixed point aluminium	BIC
		-95 °C to -80 °C	0.025 °C		BIC
		-80 °C to 248 °C	0.014 °C		BIC
		248 °C to 500 °C	0.021 °C		BIC
		500 °C to 660 °C	0.053 °C		BIC
		660 °C to 1000 °C	0.80 °C		BIC
TE 6 2	Radiation thermometers (infrared)	-35 °C to 50 °C	0.25 °C	pyrometers	BIC
		50 °C to 550 °C	0.55 °C	pyrometers	
		550 °C to 1000 °C	3.5 °C	pyrometers	
		-10 °C to 50 °C	0.40 °C	Thermal Imagers, spot calibration only	BIC
		50 °C to 550 °C	0.40 °C to 0.70 °C	Thermal Imagers, spot calibration only	
		550 °C to 1000 °C	3.5 °C	Thermal Imagers, spot calibration only	

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 013**

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024** to **01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ²	Remarks	Location
TE 6 2	Radiation thermometers (infrared)	-35 °C to -15 °C	0.35 °C	black body sources	BIC
		-15 °C to 120 °C	0.19 °C	black body sources	
		120 °C to 500 °C	0.19 °C to 0.47 °C	black body sources	
TE 9 0	Simulators/ indicators				
TE 9 1	For the purpose of resistance thermometers	-200 °C to 850 °C	0.006 °C to 0.009 °C	measuring and generating 4) based on Pt100	BIC + OS
TE 9 2	For the purpose of thermocouples	0 °C to 26 °C	0.025 °C	measuring and generating 5) thermocouple type E	BIC
		-250 °C to -200 °C	0.38 °C	measuring and generating 5)	BIC + OS
		-200 °C to -100 °C	0.25 °C	measuring and generating 5)	BIC + OS
		-100 °C to -25 °C	0.14 °C	measuring and generating 5)	BIC + OS
		-25 °C to 120 °C	0.12 °C	measuring and generating 5)	BIC + OS
		120 °C to 1000 °C	0.19 °C	measuring and generating 5)	BIC + OS
		1000 °C to 1372 °C	0.30 °C	measuring and generating 5)	BIC + OS
		1372 °C to 1767 °C	0.34 °C	measuring and generating 5)	BIC + OS
TE 13 0	Other temperature enclosures				
TE 13 1	Dry block calibrator	-95 °C to 140 °C	0.033 °C		BIC
		140 °C to 660 °C	0.033 °C to 0.053 °C		BIC
		660 °C to 1000 °C	0.56 °C		BIC
		1000 °C to 1200 °C	2.2 °C		BIC

of **Fluke Nederland B.V.**
Standaard Laboratorium

This annex is valid from: **01-05-2024 to 01-10-2027**

Replaces annex dated: **20-07-2023**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ²	Remarks	Location
TE 13 2	Thermostat baths	-80 °C to 500 °C	0.025 °C		BIC
RH 0 0	Humidity				
RH 1 3	Relative Humidity sensors	10 %rh to 70 %rh	0.37 %rh	15 °C to 50 °C	BIC
		70 %rh to 95 %rh	0.46 %rh	15 °C to 50 °C	

Remarks:

- Calibrations inside the pressure and mechanical laboratory are carried out at an ambient temperature of nominal (20 ± 2) °C, with a relative humidity of nominal (45 ± 20) %rh.
 - Calibrations inside the electrical laboratory are carried out at an ambient temperature of nominal (23 ± 1) °C, with a relative humidity of nominal (45 ± 10) %rh.
 - Temperature calibrations are carried out at an ambient temperature of nominal (23 ± 3) °C, with a relative humidity of nominal (45 ± 20) %rh.
 - The calibrations outside the electrical laboratory are carried out at an ambient temperature of nominal (23 ± 3) °C, with a relative humidity of nominal (45 ± 20) %rh.
 - Calibration on customer location are carried out at an ambient temperature of nominal (23 ± 5) °C, with a relative humidity of nominal (45 ± 20) %rh.
- 1) T_e = Trigger error for sine wave signals = $(4/f) \cdot 10^{-5} \cdot f$ (f = measured frequency).
 - 2) Generate at $T_a = (23 \pm 3)$ °C.
 - 3) T.E. = trigger error related to number of 10 MHz pulses counting during start/stop.
 - 4) Resistance thermometers based on a Pt100. For others e.g. thermistors which actually measure resistance, see calibration measurement capabilities for DC/LF quantities.
 - 5) Thermocouple measurement with or without internal reference junction compensation, actually measuring temperature dependent voltage. See calibration measurement capabilities for DC/LF quantities. Uncertainty depends on thermocouple type.