

of **Minerva Meettechniek B.V.**

This annex is valid from: **05-07-2023** to **01-01-2025**

Replaces annex dated: **04-08-2021**

Location(s) where activities are performed under accreditation

Head Office

Chrysantstraat 1
 3812 WX
 Amersfoort
 Nederland

Location	Abbreviation/ location code
Chrysantstraat 1 3812 WX Amersfoort Nederland	AM

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
MW 1 0	Mass	100 mg - 11 kg	$1,0 \cdot 10^{-5} \cdot m_c + 0,05 \text{ mg}$	Mass piece density $\geq 6400 \text{ kg/m}^3$	AM
	True Mass	100 mg - 11 kg	$1,1 \cdot 10^{-5} \cdot m + 0,05 \text{ mg}$		
PV 1 1	Absolute pressure	0 kPa - 15 kPa	$2,9 \cdot 10^{-5} \cdot p + 0,008 \text{ Pa}$	Nitrogen determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM
		5 kPa - 190 kPa	$1,8 \cdot 10^{-5} \cdot p + 0,5 \text{ Pa}$		
		25 kPa - 2,5 MPa	$2,8 \cdot 10^{-5} \cdot p + 0,5 \text{ Pa}$		

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas
 Director of Operations

of **Minerva Meettechniek B.V.**

This annex is valid from: **05-07-2023 to 01-01-2025**

Replaces annex dated: **04-08-2021**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
		50 kPa - 5 MPa	$3,0 \cdot 10^{-5} \cdot p + 0,5 \text{ Pa}$		
		300 kPa - 20 MPa	$3,1 \cdot 10^{-5} \cdot (p - p_{amb}) + 5,5 \text{ Pa}$		
		1 MPa - 70 MPa	$4,0 \cdot 10^{-5} \cdot (p - p_{amb}) + 19 \text{ Pa}$		
PV 1 2	Gauge pressure	0 kPa - 15 kPa	$3,0 \cdot 10^{-5} \cdot p_e + 0,005 \text{ Pa}$	Nitrogen determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM
		5 kPa - 190 kPa	$1,8 \cdot 10^{-5} \cdot p_e + 0,12 \text{ Pa}$		
		25 kPa - 2,5 MPa	$2,8 \cdot 10^{-5} \cdot p_e + 0,06 \text{ Pa}$		
		50 kPa - 5 MPa	$3,0 \cdot 10^{-5} \cdot p_e + 0,12 \text{ Pa}$		
		200 kPa - 20 MPa	$3,1 \cdot 10^{-5} \cdot p_e + 3,8 \text{ Pa}$		
		1 MPa - 70 MPa	$4,0 \cdot 10^{-5} \cdot p_e + 19 \text{ Pa}$		
PV 2 1	Absolute pressure	600 kPa - 50 MPa	$3,1 \cdot 10^{-5} \cdot (p - p_{amb}) + 31 \text{ Pa}$	Oil determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM
		2 MPa - 200 MPa	$4,2 \cdot 10^{-5} \cdot (p - p_{amb}) + 50 \text{ Pa}$		
		5 MPa - 500 MPa	$5,8 \cdot 10^{-5} \cdot (p - p_{amb}) + 0,1 \text{ kPa}$		
PV 2 2	Gauge pressure	500 kPa - 50 MPa	$3,1 \cdot 10^{-5} \cdot p_e + 30 \text{ Pa}$	Oil determination of effective area by means of cross-floating calibration of secondary standards and pressure devices	AM
		2 MPa - 200 MPa	$4,2 \cdot 10^{-5} \cdot p_e + 50 \text{ Pa}$		
		5 MPa - 500 MPa	$5,8 \cdot 10^{-5} \cdot p_e + 0,1 \text{ kPa}$		

of **Minerva Meettechniek B.V.**

This annex is valid from: **05-07-2023** to **01-01-2025**

Replaces annex dated: **04-08-2021**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Differential pressure on elevated line pressure	8 MPa (max. line pressure)	$1 \cdot 10^{-6} \cdot p_e + 5,6 \cdot 10^{-5} \cdot \Delta p + 13 \text{ Pa}$	Δp = Differential pressure	
PV 3 1	Under atmospheric pressure	0 ... -15 kPa	$3,0 \cdot 10^{-5} \cdot p_e + 0,005 \text{ Pa}$	Negative gauge pressure determination of effective area by means of cross-floating	AM
		-8 ... -98 kPa	$2,8 \cdot 10^{-5} \cdot p_e + 0,12 \text{ Pa}$	calibration of secondary standards and pressure devices	

Remarks:

- This annex is applicable to calibrations carried out in the own laboratory.
- The calibrations are carried out at an ambient temperature of 20 °C (nominal).
- $p_e = p - p_{amb}$; p_e is the gauge pressure, p_{amb} is the ambient pressure.
- The accreditation for mass measurements is restricted to calibrations of weights related to pressure balances.
- For a weight with a temperature of 20 °C, the conventional mass is the mass of a reference weight with a density of 8000 kg/m³, which is in balance in air with a density of 1,2 kg/m³.

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 048**

of **Minerva Meettechniek B.V.**

This annex is valid from: **05-07-2023 to 01-01-2025**

Replaces annex dated: **04-08-2021**

HCS code	Measured quantity, Range	Frequency	CMC ²	Remarks	Location
LF 1 1	DC Voltage				AM
				measuring and generating	
	0 – 20 V		$5,8 \cdot 10^{-6} \cdot U + 2 \cdot 10^{-6} \text{ V}$		
	0 – 100 V		$8,6 \cdot 10^{-6} \cdot U + 1,2 \cdot 10^{-4} \text{ V}$		
LF 2 1	DC Current				AM
	0 – 20 mA		$1,4 \cdot 10^{-5} \cdot I + 4 \cdot 10^{-5} \text{ mA}$		
	0 – 100 mA		$5,7 \cdot 10^{-5} \cdot I + 1 \cdot 10^{-3} \text{ mA}$		

Remarks:

- This annex is applicable to calibrations carried out in the own laboratory.
- The calibrations are carried out at an ambient temperature of 20 °C (nominal).

² Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, U , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".