

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 064**

of **Scott Speciality Gases Netherlands B.V.**
Calibration Laboratory

This annex is valid from: **08-02-2023** to **01-05-2025**

Replaces annex dated: **17-03-2022**

Location(s) where activities are performed under accreditation

Head Office

Takkebijsters 46- 48
 4817 BL
 Breda
 The Netherlands

Location	Abbreviation/ location code
Takkebijsters 46 – 48 4817 BL Breda The Netherlands	Br

HCS code	Measured quantity, Instrument, Measure	Range Amount fractions (mol/mol) ²⁾	CMC ¹⁾ Relative uncertainties	Remarks / Internal Instructions	Location
----------	--	--	--	---------------------------------	----------

RM 2 0 Gas mixtures

Binary mixtures
 ISO 6143 and ISO Guide 35

	CH ₄ in Air	(1·10 ⁻⁶ – 10·10 ⁻⁶) (10·10 ⁻⁶ – 100·10 ⁻⁶)	2.0 % 1.0 %		Br
	C ₃ H ₈ in Air	(10·10 ⁻⁶ – 0.33·10 ⁻²)	1.0 %	(measured as THC)	
	NO in N ₂	(1·10 ⁻⁶ – 5·10 ⁻⁶) (5·10 ⁻⁶ – 0.1·10 ⁻²)	2.0 % 1.0 %		
	NO ₂ in Air	(3·10 ⁻⁶ – 100·10 ⁻⁶) (100·10 ⁻⁶ – 0.1·10 ⁻²)	2.0 % 1.0 %		

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

of **Scott Speciality Gases Netherlands B.V.**
Calibration Laboratory

This annex is valid from: **08-02-2023** to **01-05-2025**

Replaces annex dated: **17-03-2022**

HCS code	Measured quantity, Instrument, Measure	Range Amount fractions (mol/mol) ²⁾	CMC ¹⁾ Relative uncertainties	Remarks / Internal Instructions	Location
	NO _x in N ₂	(3·10 ⁻⁶ – 5·10 ⁻⁶) (5·10 ⁻⁶ – 100·10 ⁻⁶)	2.0 % 1.0 %	(measured as NO)	Br
	NO _x in Air	(3·10 ⁻⁶ – 100·10 ⁻⁶)	3.0 %	(measured as NO ₂)	
	O ₂ in N ₂	(0.4·10 ⁻² – 1.0·10 ⁻²) (1.0·10 ⁻² – 25·10 ⁻²)	2.0 % 1.0 %		
	CO in N ₂	(150·10 ⁻⁶ – 15·10 ⁻²)	1.0 %		
	CO ₂ in N ₂	(200·10 ⁻⁶ – 20·10 ⁻²)	1.0 %		
	SO ₂ in N ₂	(50·10 ⁻⁶ – 0.5·10 ⁻²)	1.0 %		

Natural Gas Mixtures
 ISO 6143 and ISO Guide 35

	CH ₄	(70·10 ⁻² – 92.5·10 ⁻²)	0.1 %		Br
	C ₂ H ₆	(0.3·10 ⁻² – 11·10 ⁻²)	1.0 %		
	C ₃ H ₈	(0.1·10 ⁻² – 0.5·10 ⁻²) (0.5·10 ⁻² – 5·10 ⁻²)	2.0 % 1.0 %		
	iso-C ₄ H ₁₀	(0.05·10 ⁻² – 1·10 ⁻²)	1.0 %		
	<i>n</i> -C ₄ H ₁₀	(0.03·10 ⁻² – 1·10 ⁻²)	1.0 %		
	neo-C ₅ H ₁₂	(0.005·10 ⁻² – 0.01·10 ⁻²) (0.01·10 ⁻² – 0.45·10 ⁻²)	3.0 % 2.0 %		
	iso-C ₅ H ₁₂	(0.005·10 ⁻² – 0.02·10 ⁻²) (0.02·10 ⁻² – 0.5·10 ⁻²)	2.0 % 1.0 %		
	<i>n</i> -C ₅ H ₁₂	(0.005·10 ⁻² – 0.02·10 ⁻²) (0.02·10 ⁻² – 0.6·10 ⁻²)	2.0 % 1.0 %		
	<i>n</i> -C ₆ H ₁₄	(0.005·10 ⁻² – 0.02·10 ⁻²) (0.02·10 ⁻² – 0.03·10 ⁻²) (0.03·10 ⁻² – 0.4·10 ⁻²)	3.0 % 2.0 % 1.0 %		
	CO ₂	(0.05·10 ⁻² – 0.6·10 ⁻²) (0.6·10 ⁻² – 9·10 ⁻²)	3.0 % 1.0 %		
	N ₂	(0.1·10 ⁻² – 0.6·10 ⁻²) (0.6·10 ⁻² – 20·10 ⁻²)	2.0 % 1.0 %		

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 064**

of **Scott Speciality Gases Netherlands B.V.**
Calibration Laboratory

This annex is valid from: **08-02-2023 to 01-05-2025**

Replaces annex dated: **17-03-2022**

HCS code	Measured quantity, Instrument, Measure	Range Amount fractions (mol/mol) ²⁾	CMC ¹⁾ Relative uncertainties	Remarks / Internal Instructions	Location
Calculated values for Physical Properties ISO 6976 ³⁾					
	Molar mass	N ₂ < 20·10 ⁻² CO ₂ < 9·10 ⁻² C ₂ H ₆ < 11·10 ⁻² other components < 5·10 ⁻² CH ₄ < 92.5·10 ⁻²	0.1 %		Br
	Compression factor		0.1 %		
	Gross Calorific value		0.1 %		
	Relative Density		0.1 %		
	Density		0.1 %		
	Gross Wobbe Index		0.1 %		
Calibration Gas for Motor Vehicle Exhaust Emission Measurements ⁴⁾ ISO 6143 and ISO Guide 35					
	C ₃ H ₈	(72·10 ⁻⁶ – 660·10 ⁻⁶) (660·10 ⁻⁶ – 3740·10 ⁻⁶)	2.0 % 1.0 %		Br
	CO	(0.225·10 ⁻² – 5.5·10 ⁻²)	1.0 %		
	CO ₂	(2.7·10 ⁻² – 15.4·10 ⁻²)	1.0 %		
	O ₂	(0.45·10 ⁻² – 1.5·10 ⁻²) (1.5·10 ⁻² – 22.99·10 ⁻²)	2.0 % 1.0 %		
	N ₂	Balance gas			

¹⁾ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Expression of the Uncertainty of Measurement in Calibration".

²⁾ The composition of the mixtures is being certified, prepared and analysed as mole fraction. Conversion calculations are performed in conformance with ISO14912.

³⁾ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to ISO 6976.

⁴⁾ During conversion into volume fractions the following reference conditions are presumed: Temperature: 15°C and pressure: 100,0 kPa.

Calibrations are performed on permanent laboratory premises.