

Annex to declaration of accreditation (scope of accreditation)  
 Normative document: EN ISO/IEC 17025:2017  
 Registration number: **K 065**

of **Thermo Electric Instrumentation B.V.**

This annex is valid from: **30-01-2025 to 01-04-2029**

Replaces annex dated: **10-01-2024**

**Location(s) where activities are performed under accreditation**

**Head Office**

Coenecoop 71-73  
 2741 PH  
 Waddinxveen  
 The Netherlands

Location	Abbreviation/ location code
Coenecoop 71-73 2741 PH Waddinxveen The Netherlands	Wa

HCS code	Measured quantity, Range	Frequency	CMC <sup>1</sup>	Remarks	Location
LF 0 0	DC/LF Electricity				
LF 1 0	DC Voltage			measuring	Wa
	1 µV – 100 mV		$5 \cdot 10^{-6} \cdot U + 2 \mu V$		
	100 mV – 1 V		$5 \cdot 10^{-6} \cdot U + 2 \mu V$		
	1 V – 10 V		$5.8 \cdot 10^{-6} \cdot U + 1.2 \mu V$		
	DC Voltage			generating	Wa
	1 µV – 100 mV		$5 \cdot 10^{-6} \cdot U + 2.6 \mu V$		
	100 mV – 1 V		$5 \cdot 10^{-6} \cdot U + 2.6 \mu V$		

<sup>1</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Annex to declaration of accreditation (scope of accreditation)  
 Normative document: EN ISO/IEC 17025:2017  
 Registration number: **K 065**

of **Thermo Electric Instrumentation B.V.**

This annex is valid from: **30-01-2025** to **01-04-2029**

Replaces annex dated: **10-01-2024**

HCS code	Measured quantity, Range	Frequency	CMC <sup>1</sup>	Remarks	Location
	1 V – 10 V		$6 \cdot 10^{-6} \cdot U + 1.6 \mu\text{V}$		
LF 2 0	DC Current			measuring and generating	Wa
	100 $\mu\text{A}$ – 1 mA		0.2 $\mu\text{A}$		
	1 mA – 10 mA		$9 \cdot 10^{-5} \cdot I + 0.11 \mu\text{A}$		
	10 mA – 50 mA		$1 \cdot 10^{-4} \cdot I$		
LF 6 2	DC Resistance			measuring	Wa
	1 $\Omega$ – 10 $\Omega$		3.5 m $\Omega$		
	10 $\Omega$ – 100 $\Omega$		$1 \cdot 10^{-5} \cdot R + 3.4 \text{ m}\Omega$		
	100 $\Omega$ – 1 k $\Omega$		$3 \cdot 10^{-5} \cdot R + 1.4 \text{ m}\Omega$		
	1 k $\Omega$ – 10 k $\Omega$		$1.6 \cdot 10^{-5} \cdot R + 16 \text{ m}\Omega$		
	DC Resistance			generating	Wa
	1 $\Omega$ – 10 $\Omega$		5.3 m $\Omega$		
	10 $\Omega$ – 100 $\Omega$		5.3 m $\Omega$		
	100 $\Omega$ – 1 k $\Omega$		$3 \cdot 10^{-5} \cdot R + 2.4 \text{ m}\Omega$		
	1 k $\Omega$ – 10 k $\Omega$		$1.6 \cdot 10^{-5} \cdot R + 16 \text{ m}\Omega$		

of **Thermo Electric Instrumentation B.V.**

This annex is valid from: **30-01-2025 to 01-04-2029**

Replaces annex dated: **10-01-2024**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>2</sup>	Remarks	Location
TE 0 0	Temperature				Wa
	Calibration furnaces and -bathes	-196 °C to 550 °C	0.056 °C		
		550 °C to 1050 °C	0.96 °C		
TE 1 0	Resistance thermometers (Pt100)	-196 °C	0.04 °C	Boiling point of liquid nitrogen	Wa
		0 °C	0.012 °C	Ice bath	
		-80 °C to -20 °C	0.046 °C		
		-20 °C to 90 °C	0.048 °C		
		90 °C to 200 °C	0.056 °C		
		200 °C to 550 °C	0.092 °C		
		550 °C to 850 °C	0.88 °C		
TE 3 0	Thermocouples	-196 °C	0.092 °C	Boiling point of liquid nitrogen CMC based on thermocouple type E	Wa
		0 °C	0.06 °C	Ice bath CMC based on thermocouple type E	
		-80 °C to 90 °C	0.078 °C	CMC based on thermocouple type E	
		90 °C to 200 °C	0.08 °C	CMC based on thermocouple type E	
		200 °C to 550 °C	0.11 °C	CMC based on thermocouple type E	
		550 °C to 1050 °C	0.94 °C	CMC based on thermocouple type J	
		1050 °C to 1500 °C	3.4 °C	CMC based on thermocouple type R	

<sup>2</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

of **Thermo Electric Instrumentation B.V.**

This annex is valid from: **30-01-2025 to 01-04-2029**

Replaces annex dated: **10-01-2024**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>2</sup>	Remarks	Location
TE 4 0	Self Indicating thermometers				
TE 4 1	Self Indicating thermometers, sensors with transmitter (0-20) mA and (4-20) mA	-196 °C	0.04 °C	Boiling point of liquid nitrogen	Wa
		0 °C	0.01 °C	Ice bath	
		-80 °C to 90 °C	0.045 °C		
		90 °C to 200 °C	0.054 °C		
		200 °C to 550 °C	0.088 °C		
		550 °C to 1050 °C	0.94 °C		
		1050 °C to 1500 °C	3.4 °C		
TE 9 0	Simulators / indicators				
TE 9 1	For the purpose of resistance thermometers	-200 °C to 850 °C	0.008 °C to 0.038 °C	CMC based on Pt100	Wa
TE 9 2	For the purpose of thermocouples	-200 °C to 0 °C	0.1 °C to 0.052 °C	CMC based on thermocouple type E	Wa
		0 °C to 1000 °C	0.052 °C to 0.078 °C	CMC based on thermocouple type E	
		1000 °C to 1200 °C	0.096 °C	CMC based on thermocouple type E	
		1200 °C to 1372 °C	0.13 °C	CMC based on thermocouple type K	
		1372 °C to 1768 °C	0.48 °C	CMC based on thermocouple type R	
		1768 °C to 1820 °C	0.48 °C	CMC based on thermocouple type B	

Annex to declaration of accreditation (scope of accreditation)  
Normative document: EN ISO/IEC 17025:2017  
Registration number: **K 065**

of **Thermo Electric Instrumentation B.V.**

This annex is valid from: **30-01-2025** to **01-04-2029**

Replaces annex dated: **10-01-2024**

<b>HCS code</b>	<b>Measured quantity, Instrument, Measure</b>	<b>Range</b>	<b>CMC<sup>2</sup></b>	<b>Remarks</b>	<b>Location</b>
	Transmitters	-200 °C to 850 °C	0.022 °C	CMC based on Pt100 input	Wa
		-270 °C to 1372 °C	0.08 °C	CMC based on thermocouple, type K input	

**Remarks:**

- The calibrations are carried out at an ambient temperature of 23 °C nominal.
- This list with the scope of accreditations has only a relation with calibrations performed within the laboratory at Waddinxveen.