

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 070**

of **KROHNE AG**
Production facility KROHNE Altometer
(KvK number: 23039351)

This annex is valid from: **21-11-2024** to **01-01-2029**

Replaces annex dated: **29-11-2023**

Location(s) where activities are performed under accreditation

Head Office

Kerkeplaat 12
 3313 LC
 Dordrecht
 The Netherlands

Location	Abbreviation/ location code
Kerkeplaat 12 3313 LC Dordrecht The Netherlands	DO

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
FL 0 0	Flow of liquids				
FL 1 1	Liquid flow rate (water)	(0.005 – 7.0) m ³ /s	0.026 %	PSTA04; bus method reference volume > 30 m ³	DO
			0.033 %	PSTA04; bus method reference volume > 11.6 m ³	
		(0.005 – 7.0) m ³ /s	0.029 %	PSTA04; pulse method reference volume > 30 m ³	
			0.035 %	PSTA04; pulse method reference volume > 11.6 m ³	

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2017
Registration number: **K 070**

of **KROHNE AG**
Production facility KROHNE Altometer
(KvK number: 23039351)

This annex is valid from: **21-11-2024** to **01-01-2029**

Replaces annex dated: **29-11-2023**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
FL 1 1	Liquid flow rate (water)	(0.005 – 1.5) m ³ /s	0.028 %	PSTA05; bus/pulse method reference volume > 10 m ³	DO
		(0.0007 – 1.5) m ³ /s	0.11 %	PSTA05; bus/pulse method reference volume > 0.7 m ³	
		(0.0001 – 0.045) m ³ /s	0.016 %	PSTA15; bus/pulse method	

Remarks:

Calibration with PSTA04 and PSTA05 is performed with one or several volume measurements (Gravitational method)
The calibrations are carried out at an ambient temperature of nominal 20 °C.