

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 105

of **EURO-INDEX B.V. Kalibratielaboratorium**

This annex is valid from: **22-07-2021 to 01-08-2025**

Replaces annex dated: **01-07-2020**

Location(s) where activities are performed under accreditation

Head Office

Rivium 2e Straat 12
2909 LG
Capelle a/d IJssel
The Netherlands

Location	Abbreviation/ location code
Rivium 2e Straat 12 2909 LG Capelle a/d IJssel The Netherlands	C

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 0 0	DC/LF Quantities				C
LF 1 1	DC Voltage				C
	0 mV - 330 mV		$2,8 \cdot 10^{-5} \cdot U + 1,4 \mu\text{V}$	generating	
	0,33 V - 3,3 V		$1,6 \cdot 10^{-5} \cdot U + 2,8 \mu\text{V}$		
	3,3 V - 33 V		$1,7 \cdot 10^{-5} \cdot U + 28 \mu\text{V}$		
	33 V - 330 V		$2,6 \cdot 10^{-5} \cdot U + 0,21 \text{ mV}$		
	330 V - 1000 V		$2,6 \cdot 10^{-5} \cdot U + 2,1 \text{ mV}$		
LF 2 1	DC Current				C

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, U , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 105

of **EURO-INDEX B.V. Kalibratielaboratorium**This annex is valid from: **22-07-2021 to 01-08-2025**Replaces annex dated: **01-07-2020**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
	0 mA - 0,33 mA		$2,1 \cdot 10^{-4} \cdot I + 0,03 \mu\text{A}$	generating	
	0,33 mA - 3,3 mA		$1,4 \cdot 10^{-4} \cdot I + 0,07 \mu\text{A}$		
	3,3 mA - 33 mA		$1,4 \cdot 10^{-4} \cdot I + 0,4 \mu\text{A}$		
	33 mA - 330 mA		$1,4 \cdot 10^{-4} \cdot I + 3,5 \mu\text{A}$		
	330 mA - 1,1 A		$2,8 \cdot 10^{-4} \cdot I + 56 \mu\text{A}$		
	1,1 A - 3 A		$5,3 \cdot 10^{-4} \cdot I + 56 \mu\text{A}$		
	3 A - 11 A		$7,0 \cdot 10^{-4} \cdot I + 0,7 \text{ mA}$		
	11 A - 20 A		$1,4 \cdot 10^{-3} \cdot I + 2,1 \text{ mA}$		
LF 3 1	AC Voltage				C
	10 mV - 33 mV	45 Hz – 10 kHz	$2,1 \cdot 10^{-4} \cdot U + 8,4 \mu\text{V}$	generating	
		10 kHz – 20 kHz	$2,8 \cdot 10^{-4} \cdot U + 8,4 \mu\text{V}$		
		20 kHz – 50 kHz	$1,4 \cdot 10^{-3} \cdot U + 9 \mu\text{V}$		
		50 kHz – 100 kHz	$4,9 \cdot 10^{-3} \cdot U + 17 \mu\text{V}$		
	33 mV - 330 mV	10 Hz – 45 Hz	$4,2 \cdot 10^{-4} \cdot U + 12 \mu\text{V}$		
		45 Hz – 10 kHz	$2,1 \cdot 10^{-4} \cdot U + 12 \mu\text{V}$		
		10 kHz – 20 kHz	$2,3 \cdot 10^{-4} \cdot U + 12 \mu\text{V}$		
		20 kHz – 50 kHz	$4,9 \cdot 10^{-4} \cdot U + 12 \mu\text{V}$		
		50 kHz – 100 kHz	$1,2 \cdot 10^{-3} \cdot U + 45 \mu\text{V}$		
		100 kHz – 500 kHz	$2,8 \cdot 10^{-3} \cdot U + 98 \mu\text{V}$		
	0,33 V - 3,3 V	10 Hz – 45 Hz	$4,2 \cdot 10^{-4} \cdot U + 70 \mu\text{V}$		
		45 Hz – 10 kHz	$2,1 \cdot 10^{-4} \cdot U + 84 \mu\text{V}$		
		10 kHz – 20 kHz	$2,7 \cdot 10^{-4} \cdot U + 84 \mu\text{V}$		
		20 kHz – 50 kHz	$4,2 \cdot 10^{-4} \cdot U + 70 \mu\text{V}$		
		50 kHz – 100 kHz	$9,8 \cdot 10^{-4} \cdot U + 0,18 \text{ mV}$		
		100 kHz – 500 kHz	$3,4 \cdot 10^{-3} \cdot U + 0,84 \text{ mV}$		
	3,3 V - 33 V	10 Hz – 45 Hz	$4,2 \cdot 10^{-4} \cdot U + 0,91 \text{ mV}$		

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 105

of **EURO-INDEX B.V. Kalibratielaboratorium**This annex is valid from: **22-07-2021 to 01-08-2025**Replaces annex dated: **01-07-2020**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
		45 Hz – 10 kHz	$2,1 \cdot 10^{-4} \cdot U + 0,84$ mV		
		10 kHz – 20 kHz	$3,4 \cdot 10^{-4} \cdot U + 0,90$ mV		
		20 kHz – 50 kHz	$4,9 \cdot 10^{-4} \cdot U + 0,84$ mV		
		50 kHz – 100 kHz	$1,3 \cdot 10^{-3} \cdot U + 2,3$ mV		
33 V - 330 V		45 Hz – 1 kHz	$2,7 \cdot 10^{-4} \cdot U + 2,8$ mV		
		1 kHz – 10 kHz	$2,8 \cdot 10^{-4} \cdot U + 9$ mV		
		10 kHz – 20 kHz	$3,5 \cdot 10^{-4} \cdot U + 8,4$ mV		
		20 kHz – 50 kHz	$4,2 \cdot 10^{-4} \cdot U + 8,4$ mV		
		50 kHz – 100 kHz	$2,8 \cdot 10^{-3} \cdot U + 70$ mV		
330 V - 1000 V		45 Hz – 1 kHz	$4,2 \cdot 10^{-4} \cdot U + 14$ mV		
		1 kHz – 5 kHz	$3,5 \cdot 10^{-4} \cdot U + 14$ mV		
		5 kHz – 10 kHz	$4,2 \cdot 10^{-4} \cdot U + 14$ mV		
LF 4 1	AC Current				C
	0,1 mA - 0,33 mA	10 Hz – 20 Hz	$2,8 \cdot 10^{-3} \cdot I + 0,14$ µA	generating	
		20 Hz – 45 Hz	$2,1 \cdot 10^{-3} \cdot I + 0,14$ µA		
		45 Hz – 1 kHz	$1,8 \cdot 10^{-3} \cdot I + 0,14$ µA		
		1 kHz – 5 kHz	$4,2 \cdot 10^{-3} \cdot I + 0,21$ µA		
		5 kHz – 10 kHz	$1,2 \cdot 10^{-2} \cdot I + 0,28$ µA		
	0,33 mA - 3,3 mA	10 Hz – 20 Hz	$2,8 \cdot 10^{-3} \cdot I + 0,21$ µA		
		20 Hz – 45 Hz	$1,8 \cdot 10^{-3} \cdot I + 0,21$ µA		
		45 Hz – 1 kHz	$1,4 \cdot 10^{-3} \cdot I + 0,21$ µA		
		1 kHz – 5 kHz	$2,8 \cdot 10^{-3} \cdot I + 0,28$ µA		
		5 kHz – 10 kHz	$7,0 \cdot 10^{-3} \cdot I + 0,42$ µA		
	3,3 mA - 33 mA	10 Hz – 20 Hz	$2,6 \cdot 10^{-3} \cdot I + 2,8$ µA		

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 105

of **EURO-INDEX B.V. Kalibratielaboratorium**This annex is valid from: **22-07-2021 to 01-08-2025**Replaces annex dated: **01-07-2020**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 6 2	DC Resistance				C
	0 Ω - 11 Ω		$1,7 \cdot 10^{-4} \cdot R + 14 \text{ mΩ}$	generating	
	11 Ω - 110 Ω		$1,2 \cdot 10^{-4} \cdot R + 22 \text{ mΩ}$		
	110 Ω - 1100 Ω		$3,9 \cdot 10^{-5} \cdot R + 35 \text{ mΩ}$		
	1,1 kΩ - 3,3 kΩ		$3,9 \cdot 10^{-5} \cdot R + 0,35 \text{ Ω}$		
	3,3 kΩ - 11 kΩ		$3,9 \cdot 10^{-5} \cdot R + 0,21 \text{ Ω}$		
	11 kΩ - 110 kΩ		$3,9 \cdot 10^{-5} \cdot R + 2,1 \text{ Ω}$		
	110 kΩ - 1,1 MΩ		$4,5 \cdot 10^{-5} \cdot R + 21 \text{ Ω}$		
	1,1 MΩ - 3,3 MΩ		$8,4 \cdot 10^{-5} \cdot R + 0,24 \text{ kΩ}$		
	3,3 MΩ - 11 MΩ		$1,9 \cdot 10^{-4} \cdot R + 0,38 \text{ kΩ}$		
	11 MΩ - 33 MΩ		$3,5 \cdot 10^{-4} \cdot R + 3,8 \text{ kΩ}$		
	33 MΩ - 110 MΩ		$7,0 \cdot 10^{-4} \cdot R + 4,5 \text{ kΩ}$		
	110 MΩ - 330 MΩ		$4,2 \cdot 10^{-3} \cdot R + 0,14 \text{ MΩ}$		
	Resistance, decade values				
	0,1 Ω - 100 Ω		$1,5 \cdot 10^{-3} \cdot R + 5 \text{ mΩ}$	generating	
	100 Ω - 1 MΩ		$2,6 \cdot 10^{-4} \cdot R$		
	1 MΩ - 10 MΩ		$1,6 \cdot 10^{-3} \cdot R$		
	10 MΩ - 100 MΩ		$2,6 \cdot 10^{-3} \cdot R$		
	100 MΩ - 1000 MΩ		$1,6 \cdot 10^{-2} \cdot R$		
	1 GΩ - 10 GΩ		$2,1 \cdot 10^{-2} \cdot R$		
	10 GΩ - 100 GΩ		$3,1 \cdot 10^{-2} \cdot R$		
	100 GΩ - 1000 GΩ		$6,2 \cdot 10^{-2} \cdot R$		

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: **K 105**

of **EURO-INDEX B.V. Kalibratielaboratorium**

This annex is valid from: **22-07-2021** to **01-08-2025**

Replaces annex dated: **01-07-2020**

HCS code	Measured quantity, Range	Frequency	CMC ¹	Remarks	Location
LF 6 5	LF Capacitance 110 µF – 329,9 µF 330 µF - 1099,9 µF 1,1 mF - 3,3 mF		6,3·10 ⁻³ · C + 0,42 µF 6,3·10 ⁻³ · C 6,3·10 ⁻³ · C	generating Meters using charge/discharge principle. Max. rate 350 Hz	C
TF 2 1	Frequency 0,1 Hz - 2 MHz		3,5·10 ⁻⁶ · f + 7 µHz	generating	C

Remarks:

Nominal temperature for calibrations is (23 ± 3) °C

The measurements are carried out inside the own laboratory.