

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **K 138**

of **Reek Weegtechniek B.V.**

This annex is valid from: 24-08-2022 to 01-11-2025

Replaces annex dated: 21-10-2021

Location(s) where activities are performed under accreditation

Head Office

Volume 122-124
 1446 WJ
 Purmerend
 Nederland

Location	Abbreviation/ location code
Volume 122-124 1446 WJ Purmerend Nederland	PU
On-site	OS

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks ²	Location
MW	Mass and weight				
MW 1 0	Mass	1 g	0,7 mg	OIML R111, class M	PU+OS
		2 g	0,7 mg	OIML R111, class M	PU+OS
		5 g	0,7 mg	OIML R111, class M	PU+OS
		10 g	0,10 mg	OIML R111, class M	PU+OS

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

of **Reek Weegtechniek B.V.**

This annex is valid from: 24-08-2022 to 01-11-2025

Replaces annex dated: 21-10-2021

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks ²	Location
		20 g	0,16 mg	OIML R111, class M	PU+OS
		50 g	0,37 mg	OIML R111, class M	PU+OS
		100 g	0,72 mg	OIML R111, class M	PU+OS
		200 g	1,5 mg	OIML R111, class M	PU+OS
		500 g	3,6 mg	OIML R111, class M	PU+OS
		1 kg	8,6 mg	OIML R111, class M	PU+OS
		2 kg	15 mg	OIML R111, class M	PU+OS
		5 kg	36 mg	OIML R111, class M	PU+OS
		10 kg	86 mg	OIML R111, class M	PU+OS
		20 kg	0,15 g	OIML R111, class M	PU+OS
MW 2 0	Weighing instruments				
	Automatic weighing instruments	1 mg – 10 g	3,8 µg – 69 µg	Class E2	PU+OS
		10g – 100 g	69 µg – 0,1 mg	Class E2	PU+OS
		100 g – 2 kg	0,1 mg – 1,3 mg	Class E2	PU+OS
		2 kg – 40 kg	1,3 mg – 0,14 g	Class F1	PU+OS
		40 kg – 1000 kg	0,14 g – 46 g	Class M1	PU+OS
		1000 kg – 10 000 kg	46 g – 0,46 kg	Class M1	PU+OS
		10 000 kg – 80 000 kg	0,46 kg – 3,5 kg	Class M1	OS
		80 000 kg – 200 000 kg	3,5 kg – 34 kg	Class M1	OS

of **Reek Weegtechniek B.V.**

This annex is valid from: 24-08-2022 to 01-11-2025

Replaces annex dated: 21-10-2021

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks ²	Location
MW 1 2	Non-automatic weighing instruments	1 mg – 10 g	3,8 µg – 69 µg	Class E2	PU+OS
		10g – 100 g	69 µg – 0,1 mg	Class E2	PU+OS
		100 g – 2 kg	0,1 mg – 1,3 mg	Class E2	PU+OS
		2 kg – 40 kg	1,3 mg – 0,14 g	Class F1	PU+OS
		40 kg – 1000 kg	0,14 g – 46 g	Class M1	PU+OS
		1000 kg – 10 000 kg	46 g – 0,46 kg	Class M1	PU+OS
		10 000 kg – 80 000 kg	0,46 kg – 3,5 kg	Class M1	OS
		80 000 kg – 200 000 kg	3,5 kg – 34 kg	Class M1	OS

Remarks:

1 See remark CMC below page 1.

2 Noted class with weighing instruments, is the class of the used mass set. With Mass it's the class of the to be calibrated mass.