

Annex to declaration of accreditation (scope of accreditation)  
Normative document: EN ISO/IEC 17025:2017  
Registration number: **K 145**

of **Cyclertest B.V. Handelend onder Celsiuslabs  
Afdeling Kalibratie**

This annex is valid from: **20-02-2025** to **01-05-2029**

Replaces annex dated: **06-09-2023**

**Location(s) where activities are performed under accreditation**

**Head Office**

Rötscherweg 61  
6374 XW  
Landgraaf  
The Netherlands

Location	Abbreviation/ location code
Rötscherweg 61 6374 XW Landgraaf The Netherlands	LA
On site	OS

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
TE 0 0	Temperature				
TE 4 0	Self Indicating thermometers				
TE 4 1	Temperature sensors with display unit	(30 – 100) °C	0.025 °C	In liquid bath	LA
		(30 – 100) °C	0.05 °C	In multi-well dry-block calibrator	

<sup>1</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

This annex has been approved by the Board of the  
Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Annex to declaration of accreditation (scope of accreditation)  
Normative document: EN ISO/IEC 17025:2017  
Registration number: **K 145**

of **Cyclertest B.V. Handelend onder Celsiuslabs**  
**Afdeling Kalibratie**

This annex is valid from: **20-02-2025** to **01-05-2029**

Replaces annex dated: **06-09-2023**

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
		(30 – 95) °C	0.2 °C	In single-well dry-block calibrator	LA
TE 13 0	Other temperature enclosures	(30 – 95) °C	0.1 °C *	- (q)PCR cycluser * - Thermocycler * - temperature controlled dry-block incubators *	OS

\* Calibrations TE 13 0 on Location customer are performed by a multi-channel calibration, where the temperature of the thermocycler reaction block is measured simultaneously and dynamically in at least 12.5 % of the wells (or minimum 12 wells when reaction block  $\geq$  96 wells) with a measurement interval of maximum 500 ms. The calibrations TE 13 0 (q)PCR cycluser and thermocycler are performed according to ISO 20836:2021. The calibration is performed at least at 3 relevant temperature set points in the range as stated in the scope. For TE 13.0 the contributions from the device under calibration are excluded from the CMC statement.

The ambient temperature during calibrations in the own laboratory is nominal  $23 \pm 5$  °C