Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 176

## of A.P. van den Berg Ingenieursburo B.V.

This annex is valid from: 21-12-2023 to 01-02-2028 Replaces annex dated: 25-10-2023

## Location(s) where activities are performed under accreditation

## **Head Office**

Komeet 34 8448 CG Heerenveen The Netherlands

Location	Abbreviation/ location code
Komeet 34 8448 CG Heerenveen The Netherlands	Н

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
FQ 0 0	Force				Н
	Force transducer in Icone ( <i>F</i> )	(0 – 20) kN	6,3 N + 1,5·10 <sup>-3</sup> · <i>F</i>	Fin N	
	Force transducer in Icone ( <i>F</i> )	(0 – 120) kN	7,5 N + 1,7·10 <sup>-3</sup> · <i>F</i>	Fin N	
	Force transducer in Measuring head or Total load ( <i>F</i> )	(0 – 200) kN	9,5 N + 2,3·10 <sup>-3</sup> · <i>F</i>	F in N	
	Cone resistance (q <sub>c</sub> )	(0 – 150) MPa	7,5 kPa + 2,2·10 <sup>-3</sup> · <i>q</i> <sub>c</sub>	<i>q</i> c in kPa	
	Sleeve friction (f <sub>s</sub> )	(0 – 3) MPa	0,41 kPa + 2,1·10 <sup>-3</sup> ·f <sub>s</sub>	f₅ in kPa	

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Dutch Accreditation Council RvA Page 1 of 2

<sup>&</sup>lt;sup>1</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measuHrement range Measurement uncertainty, *U*, is calculated according to EA-4/02 "*Evaluation* of the Uncertainty of Measurement in Calibration".

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 176

## of A.P. van den Berg Ingenieursburo B.V.

This annex is valid from: **21-12-2023** to **01-02-2028** Replaces annex dated: **25-10-2023** 

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
PV 0 0	Pressure and vacuum				Н
PV 2 0	Pressure transducer in Icone (p)	(0 – 2) MPa	2,7 kPa + 9,1⋅10 <sup>-4</sup> ⋅ <i>p</i>	p in kPa	
	Pressure transducer in Icone (p)	(0 – 10) MPa	10,2 kPa + 1,2·10 <sup>-3</sup> · <i>p</i>	p in kPa	
	Pressure transducer in Icone (p)	(0 – 20) MPa	30,1 kPa+ 4,5·10 <sup>-4</sup> · <i>p</i>	p in kPa	
DM 0 0	Dimensional quantities				Н
DM 12 0	Clinometer in Icone	-30° to +30°	0,34°		

Dutch Accreditation Council RvA Page 2 of 2