

of **A.P. van den Berg Ingenieursburo B.V.**

This annex is valid from: **18-03-2026 to 01-02-2028**

Replaces annex dated: **03-12-2025**

Location(s) where activities are performed under accreditation

Head Office

Komeet 34
 8448 CG
 Heerenveen
 The Netherlands

Location	Abbreviation/ location code
Komeet 34 8448 CG Heerenveen The Netherlands	H

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
FQ 0 0	Force				H
	Force transducer in Icone (<i>F</i>)	(0 – 20) kN	$22,5 \text{ N} + 1,0 \cdot 10^{-3} \cdot F$	<i>F</i> in N NEN-EN-ISO 22476-1:2012	
	Force transducer in Icone (<i>F</i>)	(0 – 120) kN	$23,3 \text{ N} + 1,5 \cdot 10^{-3} \cdot F$	<i>F</i> in N NEN-EN-ISO 22476-1:2012	
	Force transducer in Measuring head or Total load (<i>F</i>)	(0 – 200) kN	$9,5 \text{ N} + 2,3 \cdot 10^{-3} \cdot F$	<i>F</i> in N NEN-EN-ISO 22476-12:2009	
	Cone resistance (<i>q_c</i>)	(0 – 150) MPa	$15,4 \text{ kPa} + 1,8 \cdot 10^{-3} \cdot q_c$	<i>q_c</i> in kPa NEN-EN-ISO 22476-1:2012	

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".

of **A.P. van den Berg Ingenieursburo B.V.**

This annex is valid from: **18-03-2026 to 01-02-2028**

Replaces annex dated: **03-12-2025**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
	Sleeve friction (f_s)	(0 – 3) MPa	$0,9 \text{ kPa} + 1,2 \cdot 10^{-3} \cdot f_s$	F in N NEN-EN-ISO 22476-1:2012	
	Force transducer in Icone (F)	(0 – 20) kN	$35,0 \text{ N} + 5,0 \cdot 10^{-3} \cdot F$	F in N NEN-EN-ISO 22476-1	
	Force transducer in Icone (F)	(0 – 120) kN	$42,0 \text{ N} + 2,5 \cdot 10^{-3} \cdot F$	F in N NEN-EN-ISO 22476-1	
	Sleeve friction for 10 cm ² Icone (f_s)	(0 – 3) MPa	$2,5 \text{ kPa} + 5,5 \cdot 10^{-3} \cdot f_s$	f_s in kPa NEN-EN-ISO 22476-1	
	Conus resistance for 10cm ² Icone (q_c)	(0 – 150) MPa	$38,0 \text{ kPa} + 3,5 \cdot 10^{-3} \cdot q_c$	q_c in kPa NEN-EN-ISO 22476-1	
PV 0 0	Pressure and vacuum				H
PV 2 0	Pressure transducer in Icone (p)	(0 – 2) MPa	$2,7 \text{ kPa} + 9,1 \cdot 10^{-4} \cdot p$	p in kPa NEN-EN-ISO 22476-1:2012	
	Pressure transducer in Icone (p)	(0 – 10) MPa	$10,2 \text{ kPa} + 1,2 \cdot 10^{-3} \cdot p$	p in kPa NEN-EN-ISO 22476-1:2012	
	Pressure transducer in Icone (p)	(0 – 20) MPa	$30,1 \text{ kPa} + 4,5 \cdot 10^{-4} \cdot p$	p in kPa NEN-EN-ISO 22476-1:2012	
	Pressure transducer in Icone (p)	(0 – 2) MPa	$2,0 \text{ kPa} + 6,0 \cdot 10^{-3} \cdot p$	p in kPa NEN-EN-ISO 22476-1	
	Pressure transducer in Icone (p)	(0 – 10) MPa	$8,0 \text{ kPa} + 5,0 \cdot 10^{-3} \cdot p$	p in kPa NEN-EN-ISO 22476-1	
	Pressure transducer in Icone (p)	(0 – 20) MPa	$12,0 \text{ kPa} + 5,0 \cdot 10^{-3} \cdot p$	p in kPa NEN-EN-ISO 22476-1	

Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2017
Registration number: **K 176**

of **A.P. van den Berg Ingenieursburo B.V.**

This annex is valid from: **18-03-2026** to **01-02-2028**

Replaces annex dated: **03-12-2025**

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
DM 0 0	Dimensional quantities				H
DM 12 0	Clinometer in Icone (I)	-30° to +30°	0,34°	/ in ° NEN-EN-ISO 22476-1:2012	
	Clinicometer in Icone (I)	-30° tot +30°	0,3° + 5,0·10 ⁻⁴ ·l	/ in ° NEN-EN-ISO 22476-1	