

Annex to declaration of accreditation (scope of accreditation)  
 Normative document: EN ISO/IEC 17025:2017  
 Registration number: **K 184**

of **Eleq Steenwijk B.V.**

This annex is valid from: **23-02-2022** to **01-02-2025**

Replaces annex dated: **21-01-2021**

**Location(s) where activities are performed under accreditation**

**Head Office**

Tukseweg 130  
 8331 LH  
 Steenwijk  
 Nederland

Location	Abbreviation/ location code
Tukseweg 130 8331 LH Steenwijk Nederland	STW

HCS code	Measured quantity, Range	Frequency	CMC <sup>1</sup>	Remarks	Location
LF 0 0	DC/LF electricity				
LF 4 0	Alternating current				
LF 4 2	AC ratio			Current Transformers	STW
	Magnitude ratio error: -2 % to +2 %	50 Hz	0.028 %	Nominal Primary current: 5 A to 3000 A (1 % - 120 %) Nominal Secondary current: 1 A; 5 A	
	Phase displacement: -120 min up to 120 min	50 Hz	1.6 min	Nominal Primary current: 5 A to 3000 A (1 % - 120 %) Nominal Secondary current: 1 A; 5 A	

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

<sup>1</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty,  $U$ , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".