Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 185

## of SenseAnywhere B.V. Calibration Laboratory

This annex is valid from: 20-10-2022 to 01-11-2026 Replaces annex dated: -

## Location(s) where activities are performed under accreditation

## **Head Office**

Bergrand 218 4707 AT Roosendaal The Netherlands

Location	Abbreviation/ location code		
Bergrand 218 4707 AT Roosendaal The Netherlands	ROO		

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
TE 0 0	Temperature				
TE 4 1	Platinum resistance thermometers with display unit	-90 °C – -80 °C	0.045 °C	Calibration in temperature controlled liquid bath	ROO
		-80 °C – -70 °C	0.036 °C		
		-70 °C – 20 °C	0.032 °C		
		20 °C – 50 °C	0.041 °C		
		50 °C – 60 °C	0.052 °C		

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Dutch Accreditation Council RvA Page 1 of 2

<sup>&</sup>lt;sup>1</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "*Evaluation* of the Uncertainty of Measurement in Calibration".

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 185

## of SenseAnywhere B.V. Calibration Laboratory

This annex is valid from: 20-10-2022 to 01-11-2026

Replaces annex dated: -

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
TE 4 1	Temperature sensors with display unit (digital system thermometer or dataloggers)	-40 °C – -30 °C	0.042 °C	Calibration in temperature controlled liquid bath	ROO
		-30 °C – 50 °C	0.041 °C		
		50 °C – 60 °C	0.052 °C		
RH 0 0	Humidity				
RH 1 3	Relative humidity sensors (Air temperature 23 °C – 30 °C)	10 %rh – 50 %rh 50 %rh – 70 %rh 70 %rh – 90 %rh	1.7 %rh 2.0 %rh 3.3 %rh	Calibration in <i>RH</i> and temperature controlled climate chamber	ROO

Dutch Accreditation Council RvA Page 2 of 2