Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2017
Registration number: **K 187**

of **Raditeq B.V.**

This annex is valid from: **22-05-2025** to **01-06-2029** Replaces annex dated: **n.v.t.**

Location(s) where activities are performed under accreditation

Head Office

Vijzelmolenlaan 3
3447 GX
Woerden
The Netherlands

Location	Abbreviation/ location code
Vijzelmolenlaan 3 3447 GX Woerden The Netherlands	WDN

HCS code	Measured quantity, Range	Frequency	CMC[1]	Remarks	Location
RF 0 0	High frequency electricity				
RF 5 0	Electrical / magnetic field quantities / EMC / EMI / TEM cell				WDN
	Electrical Field Strength 20 V/m	9 kHz – 950 MHz	1.4 dB	GTEM-CEL, transfer standard method EN61000-4-3	
		1000 MHz – 40 GHz	1.4 dB	Anechoic chamber, transfer standard method EN61000-4-3	
	Electrical Field Linearity 6 V/m +/- 6 dB 20 V/m +/- 6 dB	1 MHz and 100 MHz	0.15 dB	GTEM-CEL, EN 61000-4-3	

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

[1] Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "*Evaluation* of the Uncertainty of Measurement in Calibration".