Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: L 402

Witteveen+Bos Raadgevende ingenieurs b.v. of luchtemissies, industrielawaai en compliance

This annex is valid from: 10-04-2024 to 01-06-2027 Replaces annex dated: 31-01-2024

Location(s) where activities are performed under accreditation

Head Office Leeuwenbrug 8 7411 TJ Deventer The Netherlands Abbreviation/ location code Location Hanzeweg 45 D 7418 AV Deventer The Netherlands Mobile Location MoLo No. Material or Type of activity¹ Internal reference number Location product Sampling **Cluster: Other Organic** Emitted air. Sampling for the determination of the LM-WV-06 D a. smoke, process content of aromatic, aliphatic and NPR-CEN/TS 13649 chlorinated hydrocarbons and and exhaust gases vinylchloride; adsorption tubes (associated test is carried out structurally by another accredited body)

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Dutch Accreditation Council RvA Page 1 of 2

¹ If there is a referral to a code starting with NAW, NAP, EA or IAF, this concerns a scheme mentioned on RvA-BR010-list. If no date or version number is mentioned for a normative document, the accreditation concerns the most current version of the document or scheme

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: L 402

of Witteveen+Bos Raadgevende ingenieurs b.v. luchtemissies, industrielawaai en compliance

This annex is valid from: 10-04-2024 to 01-06-2027 Replaces annex dated: 31-01-2024

No.	Material or product	Type of activity ¹	Internal reference number	Location
Analysis of the component odour within the framework of NTA 9065				
1.	Air and (process) gases	Determination of the odour concentration by using dynamic olfactometry	LM-WV-02 NEN-EN 13725:2003	D
		Odour/olfactometry in the framewor	rk of NTA 9065	
2.	Air and (process) gases	Determination of odour emissions; method for gas outlets, hood method (including Lindvall hood method) or leeward method, with application of the lung method or the dilution method (including related sampling)	LM-WV-05 in house method (NEN-ISO 10396:1999) NEN-EN 15259	D,MoLo
		Emission measuremen	ts	1
Cluster: Physical parameters				
3.	Emitted air, smoke, process and exhaust gases	Determination of the waste gas characteristics: flow rate; differential pressure measurement	LM-WV-04 ISO 10780, NEN-EN-ISO 16911-1	D,MoLo
4.	Emitted air, smoke, process and exhaust gases	Determination of the water vapor content (in pipes); gravimetry	LM-WV-04 NEN-EN 14790	D,MoLo
		Cluster: Gaseous (in)orga	anic	1
5.	Emitted air, smoke, process and exhaust gases	Determination of the oxygen (O ₂) content; paramagnetism (including associated sampling)	LM-WV-12 NEN-EN 14789 (sampling NEN-EN 15259)	D,MoLo
6.	Emitted air, smoke, process and exhaust gases	Determination of the C _x H _y content; FID (including associated sampling)	LM-WV-12 NEN-EN 12619 (sampling NEN-EN 15259)	D,MoLo

Dutch Accreditation Council RvA Page 2 of 2