

Annex to declaration of accreditation (scope of accreditation)
Normative document: EN ISO/IEC 17025:2017
Registration number: **L 469**

of **Kiwa Technology B.V.**
Analytisch Chemisch Laboratorium (ACL) Laboratorium Materialen (PSML)

This annex is valid from: **13-09-2023** to **01-05-2027**

Replaces annex dated: **15-03-2023**

Location(s) where activities are performed under accreditation

Head Office

Wilmersdorf 50
7327 AC
Apeldoorn
The Netherlands

Location	Abbreviation/ location code
Wilmersdorf 50 7327 AC Apeldoorn The Netherlands	AP

This annex has been approved by the Board of the
Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Annex to declaration of accreditation (scope of accreditation)
 Normative document: EN ISO/IEC 17025:2017
 Registration number: **L 469**

of **Kiwa Technology B.V.**
Analytisch Chemisch Laboratorium (ACL) Laboratorium Materialen (PSML)

This annex is valid from: **13-09-2023** to **01-05-2027**

Replaces annex dated: **15-03-2023**

No.	Material or product	Type of activity ¹	Internal reference number	Location
Natural gas / other flammable gases				
1.	Natural gas	<p>Determination of the natural gas components: Helium, Argon, Hydrogen, Oxygen, Nitrogen, Carbon Oxides, Methane, Ethane, Propane, Iso-Butane, n-Butane, Neo-Pentane, Iso-Pentane, n-Pentane, 2,2-dimethylbutane, Cyclopentane + 2,3-dimethylbutane, 2-methylpentane, n-hexane, 3-Methylpentane, C7 Hydrocarbons, C8 Hydrocarbons, Benzene, Toluene using GC-TCD/FID</p> <p>Determination of the lower and upper calorific value, the density, the relative density, compressibility and the Wobbe index based on the composition.</p>	ISO 6974 ISO 6976	AP
2.	Other flammable gases ²	<p>Determination of the components of other flammable gases: Helium, Argon, Hydrogen, Oxygen, Nitrogen, Carbon Oxides, Methane, Ethane, Propane, Iso-Butane, n-Butane, Neo-Pentane, Iso-Pentane, n-Pentane, 2,2-dimethylbutane, Cyclopentane + 2,3-dimethylbutane, 2-methylpentane, n-hexane, 3-Methylpentane, C7 Hydrocarbons, C8 Hydrocarbons, Benzene, Toluene using GC-TCD/FID</p> <p>Determination of the lower and upper calorific value, the density, the relative density, compressibility and the Wobbe index based on the composition.</p>	ISO 6974 ISO 6976	AP

¹ If there is a referral to a code starting with NAW, NAP, EA or IAF, this concerns a scheme mentioned on [RvA-BR010-lijst](#).

If no date or version number is mentioned for a normative document, the accreditation concerns the most current version of the document or scheme.

of **Kiwa Technology B.V.**
Analytisch Chemisch Laboratorium (ACL) Laboratorium Materialen (PSML)

This annex is valid from: **13-09-2023** to **01-05-2027**

Replaces annex dated: **15-03-2023**

No.	Material or product	Type of activity ¹	Internal reference number	Location
Asphalt tests				
3.	Dichloormethane extract	Determination of the content of polycyclic aromatic hydrocarbons (PAH10); GC-MS Naphthalene, phenanthrene, anthracene, fluoranthene, benzo(a)anthracene, chrysene, benzo(k)fluoranthene, benzo(a)pyrene, indeno(1,2,3-cd)pyrene, benzo(g,h,i)perylene.	WI 800 en WI 801 in-house method	AP
Plastics (physical tests)				
4.	Polyethylene	Determination of the Strain Hardening Modulus (resistance to slow crack growth); tensometer	ISO 18488	AP
5.		Determination of environmental stress cracking; full-notch creep test (FNCT)	NEN-ISO 16770	AP
6.		Determination of resistance to crack propagation; test method for slow crack growth on notched pipes	ISO 13479	AP
7.		Determination of resistance to slow crack growth under cycling loading; Cracked Round Bar test method	ISO 18489	AP